

Laboratory Accreditation

An Overview

Argie P. Leach MHS, MT(ASCP)SH



Objectives



- Accrediting agencies
- CAP role in accreditation process
- CAP inspections
- Checklists
- Everyone's participation

2

History of Improved Health Care

Over 150 years ago



Ignaz Semmelweis
Pioneer of antiseptic procedures



Florence Nightingale
Founder of Modern Nursing

- 1846 – Ignaz Semmelweis
 - Assistant Professor at John Hopkins
 - Hand washing
 - required all his medical students to wash their hands with chlorine solution before doing OB exam. The students routinely switched from doing autopsy to giving birth causing 18% of women giving birth to die from puerperal fever (childbed fever)
- 1854 – Florence Nightingale
 - Public Health Pioneer saw that poor hospital sanitation resulted in a increase in fatalities among the wounded soldiers in the Crimean War
 - Developed practices which are still in existence today
 - Sanitizing surgical tools
 - Changing bed linens regularly
 - Promoter of good nutrition and clean air

Hand washing alone can prevent 40% of hospital infections!

3

Hospital Reform Timeline



COLLEGE of AMERICAN
PATHOLOGISTS



The Joint
Commission

- 1847 American Medical Association (AMA) – formed to standardize medical education
- 1917 American College of Surgeon (ACS) proposed the "End Result System" for the Hospital Standardization Program which set minimum standards for hospitals
 - In 1918 began implementing standards to inspect hospitals – only 45% met the minimum requirements
- 1951 The Joint Commission on Accreditation of Hospitals was formed by American College of Surgeons, American College of Physicians, the American Hospital Association, the American Medical Association and the Canadian Medical Association
- 1962 College of American Pathologists implemented a Laboratory Accreditation Program
- 1991 national attention was finally gained for hospital reform when it was reported that 98,000 people die yearly in hospitals due to preventable medical error.
 - To Err is Human: Building a Safer Health System

4

Agencies

Regulating Laboratories



- CAP – College of American Pathologists
- TJC – The Joint Commission
- COLA – Commission of Office of Laboratory Accreditation
- OSHA – Occupational Safety and Health Administration
- FDA – Federal Food Administration
- AABB – American Association of Blood Bankers
- OIG – Office of Inspector General
 - Oversees Billing Fraud
- OCR – Office of Civil Rights
 - Enforces HIPAA

5

CLIA

Clinical Laboratory Improvements Amendment



- 1967 Congress passes 1st CLIA Law
 - To regulate costs and quality of health care
 - Only required hospitals and large clinical laboratory
- Impetus for CLIA '88
 - Pap Mills turning out inaccurate results
 - 1987 Series of articles which appeared in Wall Street Journal
 - Response of the public furor from one or more deaths due to False-Negative PAP smear reading
 - Congress passed law 1988
- CLIA '88 purpose is to ensure that all laboratory testing is done accurately and according to good quality standards
 - Requires employee training and competency
 - Must assess the competency of all testing personnel who handle human specimens
 - Established the requirements for performance and documentation of initial training and on going competency

6

CLIA

Clinical Laboratory Improvements Amendment Continued



- 1992 Centers for Medicare and Medicaid Services (CMS), published regulations to enforce the CLIA law
- Regulates all testing on humans for health purposes
 - Labs which provide information for diagnosis, prevention or treatment of diseases
 - Ensures accurate and reliable testing regardless of the lab and its location
- CLIA regulations place overall responsibility on the laboratory director
- CLIA '88 regulations are based on complexity of the testing the more complex the test the more stringent the standards for quality
- VA labs, research labs and forensic labs are exempted
- CMS enforces CLIA

7

CMS

Centers for Medicare and Medicaid Services



- Federal Agency with the US department of Health and Human Services (DHHS) that administers Medicare
- Partially funds Medicaid
- Sets the Standards
- Supports CLIA in partnership with CDC and FDA
- Must follow CMS regulations or you cannot bill federally funded programs
- Gives deemed authority to CAP and TJC for inspecting and accrediting labs
- CMS inspects if compliant filed or randomly selects labs
- Formerly known as Health Care Financing Administration (HCFA)

8

TJC

The Joint Commission



- 1979 Started evaluating laboratories
- 1995 TJC was deemed by CMS to certify labs under Clinical Laboratory Improvements Amendments (CLIA '88) requirements
- Laboratory standard set by TJC include wide variety of labs
 - Hospitals
 - Clinics
 - Home care facilities
 - Reference labs
- Labs that are accredited by CAP do not have to be re-inspected when the hospital the lab is located in is inspected by the Joint Commission
- 2002 Joint Commission establishes the National Patient Safety Goals and Speak Up campaign™
- 2003 Universal Protocol for preventing Wrong Site, Wrong Procedure, Wrong Person Surgery™

9

Laboratory Accreditation Program

College of American Pathologist



- Inspected and accredited laboratories since 1962
- Initially only hospital and private pathology labs were inspected
- Now inspects
 - University labs
 - Commercial reference labs
 - Physicians offices

10

Laboratory Accreditation Program

College of American Pathologists (CAP)



- Program to improve the quality of Clinical Laboratory Services
 - Voluntary participation
 - Peer Review
 - Education
 - Compliance with performance standards
- 1967 CAP was deemed by Health Care Financing Administration (HCFA) now CMS as a certifying accreditation agency which meets or exceeds CLIA standards.
- Initially approved as an accrediting agency by CMS for 3 years in 1995
- 2015 CMS renewed CAP status as an accrediting agency for 6 more years

11

OSHA

Occupational Safety and Health Administration

- Main federal agency which enforces workplace health and safety legislation for workers
- Regulations include
 - Blood Borne Pathogen
 - 1985 introduced Universal Precautions
 - 1996 became Standard Precautions to be used on all patients
 - Chemical exposure
 - Ergonomic injuries
 - Fire and Electrical Safety
- Prompts inspection:
 - Reported imminent danger in workplace
 - Fatalities or catastrophes
 - Employee complaint
 - Referral from other agencies
 - Targeted facilities
 - High hazard industries

Source: www.OSHA.gov

Criminal and Civil penalties against an organization or individuals if laws are not followed.


12

Lab Accreditations



13

Accreditation



14

University Medical Center New Orleans

Clinical Laboratory

Main Lab



Blood Gases



15

Standards for Laboratory Accreditation

College of American Pathologist

Core principles for CAP Laboratory Accreditation Program:

- **Standard I – Director and Personnel**
 - Medical Significance, Interpretation, Consultations, Correlation of Data, Quality Management, Personnel, Selection of Equipment, Delegation of functions and implementation of a safe environment.
- **Standard II – Physical Resources**
 - Space, instrumentation, data processing, supplies, environment to permit effective performance of personnel
- **Standard III – Quality Management**
 - Performance improvement, QC, Instrument maintenance, Proficiency testing, Clinical relevance
- **Standard IV – Administrative Requirements**
 - Eligibility for participation in Laboratory Accreditation Program
 - Pathology and Clinical laboratory must submit to a complete periodic on-site inspection and self inspections
 - Laboratory and laboratory director must cooperate with the Commission on Laboratory Accreditation.
 - Each accredited lab must comply with the Terms of Accreditation listed in the official notice of accreditation sent to the lab by CAP.

Source: www.cap.org

16

Inspections

- Every two years
 - Anniversary date onsite inspection
- Self inspections on the off year
 - Anniversary date
 - Administered by lab
 - Deficiencies are noted
 - Should mimic actual inspection

17

Self Inspections



- "Treat Self Inspections as a real event"
- Mock inspection
- Identify and fix potential noncompliance which could result in deficiencies in a real inspection
- Prepare –
 - Review last inspections deficiencies and make sure the lab is doing what they say they are doing
 - Review all new checklist questions
- Conduct –
 - Should be performed in one day
 - Involve bench techs in process
- Improve –
 - Summation should also be held after the inspection is completed
 - Deficiencies should be reviewed and corrective action implemented
 - Documentation should be available for next on-site review.

Source: www.cap.org

18

On-site Inspection Process



- Peer Review
- Avoid conflict of interest
 - Not having former colleagues or competitors inspect each other
- Team assigned by CAP
- Uses algorithm to match comparable labs
- CAP matches team with comparable lab
 - Test Volumes
 - University Lab Vs Reference Lab

Source: www.CAP.org

19

Ready or Not... Inspection Process begins

- 7 months before inspection window opens
 - Welcome packet in email with application link
 - Application and Organizational Profile must be completed online within 3 months
 - Application includes the following:
 - Lab Director
 - Administration
 - Staff List
 - Blackout Dates
 - Testing volumes
 - Activity Menu
 - Testing performed by each lab
 - Travel accommodations
 - Lab specific information
- 3 months before anniversary date
 - CAP Custom Lab Specific Checklist sent and available online
 - Lab reviews checklist
 - Lab sections are responsible for these checklists:
 - General, Common and Lab Specific Checklist

Source: www.CAP.org

20

Online Application

Source: www.CAP.org

21

New CAP Roster



- Personnel to include on CAP Roster:
 - High complexity testing
 - Non waived or moderately complex testing
 - Physicians who do PPM
 - Grossing Techs
 - Pathology Staff and Residents
 - Cytotechnologists
- Personnel who do not need to be included:
 - Waived Testing personnel
 - Phlebotomist
 - Histology Techs

Source: www.CAP.org

22

CLIA Roles




- Technical Supervisors
- General Supervisors
- Technical Consultant
- Laboratory Director
- All Staff Pathologists
- All personnel must be included in Staff List to be included in CLIA Role

Source: www.CAP.org

23

Non Laboratory Testing Personnel




Provider Performed Microscopy - Physicians
Nurses who perform non-waived testing

- Provider Performed Microscopy
 - Physicians
- Nurses or Techs who perform non-waived test
 - Respiratory Techs
 - Nurses who do competency evaluations
 - Waived testing personnel are not required

Source: www.CAP.org

24


Online Attachments



- Equipment List
- Organizational Chart
- Laboratory Director CV
- Personnel Roster – New process 2018
 - Names of all testing personnel
 - Template can be loaded only once
 - Updated every 6 months
 - Updated online
 - After initial template loaded all other testing personnel must be entered manually
- Level of testing complexity
- Licensure

25

CAP Profiles



- All supervisors or direct participants in the CAP inspection must create a profile on the CAP website.
- This process is a requirement even if you are faxing in application
- Personnel forms are no longer included in with the CAP application packet.

26

Laboratory Data



- Print out of all data which was entered in Organization Profile submitted to CAP
- Before submitting to CAP we are able to generate a PDF of all the data we provided.

27


CAP Checklist



- First checklist compiled in 1965
- Helps lab focus on the important criteria
- Fulfill all the requirements set by CMS
- Updated annually
- Types of Checklist
 - General
 - All Common
 - Lab Specific
- Formats for Checklist
 - PDF or EXCEL
 - Master and Custom


28

Checklist Components



29

Evidence of Compliance



- Assist the lab in preparing for the inspection and in managing ongoing compliance
- Helps the inspector and laboratory consistently understand the requirements
- Provides examples of acceptable documentation

30

Evidence of Compliance

Filter with Excel

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	32	33	34	35	36	37	38	39	40	41	42	43	44	45	46	47	48	49	50	51	52	53	54	55	56	57	58	59	60	61	62	63	64	65	66	67	68	69	70	71	72	73	74	75	76	77	78	79	80	81	82	83	84	85	86	87	88	89	90	91	92	93	94	95	96	97	98	99	100
1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	32	33	34	35	36	37	38	39	40	41	42	43	44	45	46	47	48	49	50	51	52	53	54	55	56	57	58	59	60	61	62	63	64	65	66	67	68	69	70	71	72	73	74	75	76	77	78	79	80	81	82	83	84	85	86	87	88	89	90	91	92	93	94	95	96	97	98	99	100

31

Inspector Instructions

Read, Observe, Ask, Discover


	Inspector Instructions: Read, Observe, Ask, Discover Read: Review the laboratory's quality management system (QMS) documentation, including the laboratory's accreditation body (AB) requirements, the laboratory's policies and procedures, and the laboratory's quality control (QC) records. Observe: Observe the laboratory's operations, including the laboratory's personnel, equipment, and facilities. Ask: Ask the laboratory personnel for information about the laboratory's operations, including the laboratory's policies and procedures, and the laboratory's QC records. Discover: Discover the laboratory's strengths and weaknesses, including the laboratory's accreditation body (AB) requirements, the laboratory's policies and procedures, and the laboratory's QC records.
--	--

Similar to Tracer Methodology used by The Joint Commission

32

Tracer Methodology

Follow the Specimen



- Begin tracer with test result
- Follow entire testing process
- Includes Pre-analytical, analytical, post analytical process
- Pre-analytical
 - Ordering
 - Collecting procedures
 - Identifying patient
 - Labeling in front of patient
 - Transporting of specimen
- Analytical
 - QC
 - Proficiency testing results for the analyte in question
- Post analytical
 - Computer print out
 - Reporting of panic values
- Personnel involved
 - Review Delegations
 - Review Competencies

Source: www.CAP.org

33

CAP General Checklist



Section	Item	Yes	No	Not Applicable
Quality Management	1. Quality Management System (QMS) is in place and functioning.			
Specimen Collection, Data Handling and Reporting	2. Specimen Collection, Data Handling and Reporting is in place and functioning.			
Quality of Water and Glassware	3. Quality of Water and Glassware is in place and functioning.			
Laboratory Computer Systems	4. Laboratory Computer Systems is in place and functioning.			
System Maintenance	5. System Maintenance is in place and functioning.			
Personnel	6. Personnel is in place and functioning.			
Physical Facilities	7. Physical Facilities is in place and functioning.			
Lab Safety	8. Lab Safety is in place and functioning.			
Numbering system – GEN.00000	9. Numbering system – GEN.00000 is in place and functioning.			

Laboratory Administration is responsible for the Laboratory General Checklist compliance. Lab administrators may delegate duties to other laboratory areas.

33

CAP Common Checklist



Source: www.CAP.org

Laboratory Areas must review All Common for lab specific criteria.

- Introduced 2011
- Consolidated items that were similar across disciplines
 - Proficiency Testing
 - Procedure Manual
 - Results Reporting
 - Reagents
 - Instruments and Equipment
 - Maintenance and Function Checks
 - Thermometers
 - Temperature Dependent Instruments, Equipment and Environments
 - Test Method Validation
 - Method Performance Specifications
 - Reference Intervals
 - Individualized Quality Control Plan (IQCP)
 - Numbering system – COM.00000

35

Individualized Quality Control Plan (IQCP)

Section	Item	Yes	No	Not Applicable
Quality Management	1. Quality Management System (QMS) is in place and functioning.			
Specimen Collection, Data Handling and Reporting	2. Specimen Collection, Data Handling and Reporting is in place and functioning.			
Quality of Water and Glassware	3. Quality of Water and Glassware is in place and functioning.			
Laboratory Computer Systems	4. Laboratory Computer Systems is in place and functioning.			
System Maintenance	5. System Maintenance is in place and functioning.			
Personnel	6. Personnel is in place and functioning.			
Physical Facilities	7. Physical Facilities is in place and functioning.			
Lab Safety	8. Lab Safety is in place and functioning.			
Numbering system – GEN.00000	9. Numbering system – GEN.00000 is in place and functioning.			

- It was developed for non-waived testing in which QC is not performed on a daily basis
- Introduced in 2014 required implementation in 2016
- Customary QC plan developed for the laboratory specific to your environment to provide reliable and accurate patient testing
- Labs must continue to follow manufacturer's instructions for QC
 - A lab may not implement an IQCP that allows for quality control to be performed less frequently than indicated by the manufacturer
- Does not apply to waived testing
- If a lab does not follow IQCP guidelines it must run at least two levels of external QC each day of patient testing.

36

IQCP Program

Allows continued use of Electronic Quality Control

- Risk Assessment
 - Identifies and evaluates potential failures in the testing process and must address all 5 areas listed through all 3 phases of testing
 - Specimen
 - Environment
 - Reagents
 - Test system
 - Testing personnel
- Quality Control Plan
 - Ensures accuracy and reliability of test results
 - Internal Controls
 - Electronic Controls
 - Calibration
 - Training and Competency
- Quality Assessment
 - Continuous process of monitoring effectiveness of QC program
 - QC review
 - PT review
 - Compliant reports

37

Laboratory Specific Checklist



- Collection requirements
 - Anticoagulants
 - Rejection criteria
- Procedural Manual
- Calibration Procedures
- Quality Control
- Validation Procedures
- Corrective Actions

Download checklist in EXCEL Format easier to concentrate on new and questions with evidence of compliance

38

Day of Inspection



- Team Arrival
 - Arrival time usually 7:30-8:00
 - Team is met at information desk
- Opening Conference
 - Team leader introduces inspectors to lab
 - Lab director introduces self and all lab personnel present to team
 - Lab supervisors are paired with inspectors
 - Inspectors are given a tour of the lab which lasts 15-20 minutes
- Inspection
 - Review of lab manuals / documents requested
 - Inspector uses R.O.A.D and interacts with lab staff
 - Refer to the lab roster to determine staff personnel files for review
 - Interacts with staff with a FOCUS on lab process, lab safety, QCC, Proficiency testing procedures, and corrective actions
 - Inspector should discuss all deficiencies as they are found with lab supervisor
- Summation
 - Team thanks the lab for hosting them
 - Present deficiencies and how to improve

39

Summation



- Team will gather all information
 - Deficiencies
 - Recommendations
- Take all recommendations and deficiencies seriously
- Address all issues

40


Deficiencies

- Must be submitted within 30 days of inspection
- All deficiencies must be sent together
- Contact CAP if you have any questions about response or challenging
- Be vigilant and look for any responses from CAP for additional information about deficiencies responses submitted

Source: www.CAP.org

41

Most Commonly Cited Deficiencies



Source: CAP

42

Most Commonly Cited Deficiencies

on CMS radar



Source: www.cap.org

CAP is very strict about these categories – are high priority standards set by CMS – accreditation can be pulled if lab is not in compliance!

- Personnel Qualifications
 - Delegation
 - Continuing Ed
 - Licensure
- Competency
 - High Complexity Testing
 - 6 levels of competency
 - All Test Systems
 - Waived
 - 2 levels of competency
- Proficiency Testing
 - Stringent rules

43

Director Responsibilities

WHAT ARE THE RESPONSIBILITIES FOR WHICH I MUST MAINTAIN DIRECT OVERSIGHT AND CANNOT DELEGATE TO OTHERS?

- As laboratory director, you must ensure that:
 - testing systems in the laboratory provide quality services in all aspects of test performance, i.e., the preanalytic, analytic, and postanalytic phases of testing, and are appropriate for your patient population;
 - physical and environmental conditions of the laboratory are adequate and appropriate for the testing performed;
 - the environment for employees is safe from physical, chemical, and biological hazards and safety and biohazard requirements are followed;
 - a general supervisor (high complexity testing) is available to provide day-to-day supervision of all testing personnel and reporting of test results as well as provide on-site supervision for specific, manually qualified testing personnel when they are performing high complexity testing;
 - sufficient numbers of appropriately educated, experienced, and/or trained personnel who provide appropriate consultation, properly supervise, and accurately perform tests and report test results in accordance with the written duties and responsibilities specified by you, as employed by the laboratory;
 - new test procedures are reviewed, included in the procedures manual and followed by personnel; and
 - each employee's responsibilities and duties are specified in writing.


<https://www.cms.gov/Regulations-and-Guidance/regulations/CLIA/Downloads/directors7.pdf>

- Testing systems provide quality
 - Preanalytic, analytic and postanalytic phases of testing
- Physical environment
- Safety of employees
- A general supervisor is available to provide day to day supervision
- Trained personnel who can accurately perform and report out test
- New procedures are reviewed
- Each employee responsibilities and duties are specified in WRITING

44

Delegations

CLIA Laboratory Director



- Laboratory Director must delegate in writing the test system each testing personnel can perform
- Personnel Must have on file competency to mirror the delegations assigned
- CAP Inspectors may compare delegations with competency assessments and test system assignments
- If job assignments have changed then Delegation MUST be updated to reflect the duties assigned

Test system this tech is competent to perform must be listed in the delegation


(15) Specify, in writing, the responsibilities and duties of each consultant and each supervisor, as well as each person engaged in the performance of the preanalytic, analytic, and postanalytic phases of testing, that identifies which examinations and procedures each individual is authorized to perform, whether supervision is required for personnel performing, test performance to result together and whether supervisor or director review is required. Specify required patient test results. <https://www.cms.gov/Regulations-and-Guidance/regulations/CLIA/Downloads/directors7.pdf>

45

Personnel Qualifications

Common Deficiency

- CMS requires that every employee has documentation of all his or her education and experience on file. "If one person – and you might have a lab of 50 to 60 people – is missing something, it's an automatic citation."
- CAP is less stringent, but now is under much scrutiny by CMS to make sure all labs comply with this ruling.
 - Cleveland Clinic
 - CMS Validation Inspection
- CMS requires license (for technologists only, nurses, respiratory therapist need transcript/diploma) instead of diploma/transcripts in states where licensure is a requirement
- CAP requires the lab director to delegate in writing the responsibilities of each person involved in testing
- CAP addresses this issue in the General Checklist under Personnel



Make sure your personnel file is up to date

46

Reagent Labeling and Storage

Common Deficiency



- Properly Stored
- Properly labeled
- Temperature and Humidity Monitored
- Used within expiration dates

47

Procedure Manual

Common Deficiency



- Procedure same as practice
- All personnel have signed and it is available
- Must be reviewed biannually by director or designee
- Electronic downtime backup

48

Attestation Page

Common Deficiency

****REVISED** 08/17/2016**

COM-01400 PT Attestation Statement Phase II

The proficiency testing attestation statement is signed by the laboratory director or qualified designee and all individuals involved in the testing process.

NOTE: Physical signatures must appear on a paper version of the attestation form. A listing of typed names on the attestation statement does not meet the intent of the requirement. The signature of the laboratory director or designee need not be obtained prior to reporting results to the proficiency testing provider.

Designees must be qualified through education and experience to meet the defined regulatory requirements associated with the complexity of the testing as defined in the Personnel section of the Laboratory General Checklist.

- For high complexity testing, it may be delegated to an individual meeting the qualifications of a technical supervisor or section director (GEN.53400). For the specialties of Histocompatibility, Cytogenetics, and Transfusion Medicine, refer to specific requirements for the qualifications of section director/technical supervisors in the associated checklists (HSC-40000, CYG-50000, and TRM-50050).
- For moderate complexity testing, it may be delegated to an individual meeting the qualifications of a technical consultant (GEN.53625).

Evidence of Compliance:

- ✓ Appropriately signed attestation statement from submitted PT result forms

REFERENCES:

1) Department of Health and Human Services, Centers for Medicare and Medicaid Services. Clinical laboratory improvement amendments of 1988. Final rule. Fed Register. 1992;57(28):7148-142CFR493.801(b)(1)(i)

Source: www.CAP.org

49

Proficiency Testing

Common Deficiency

Source: www.CAP.org

1967 Proficiency Testing participation becomes a requirement for accreditation

- Samples MUST be performed with routine laboratory workload
- Samples MUST be analyzed by personnel who routinely test patient samples
- Samples Must be rotated among all testing personnel and NOT just dedicated operators or instruments
- Samples MUST be run like patient sample
 - No duplicate testing
 - Lab personnel MUST notify CMS if it receives a PT testing request from another laboratory
- Personnel are STRICTLY prohibited from communicating with other testing labs about PT samples

50

Document Control

Common Deficiency



<https://www.medialab.org>

- All Documents must be included
 - Forms
 - Job aids or cheat sheets
- Make sure that only current documents or forms are in use
- All current documents are properly approved

51

Activity Menu

Common Deficiency

Source: www.CAP.org

- Make sure all testing in included in the Activity Menu
- If stopped testing – Remove
- If Started testing – Insert
- Make sure CAP is aware of all changes

52

Competency

Common Deficiency

Waived Testing only requires 2 levels of competency

Initials and dates the performance level of competency

- Non-waived
 - High Complexity
 - Moderately Complex
- Assessment Program
 - Defined set of skills/knowledge of job
 - Critical for levels of acceptance
 - Process of reevaluation
 - Six components of competency
- All 6 levels of competency must be assessed annually for each test system
 - Direct Observation
 - Recording/Reporting of Results
 - Review of QC
 - Direct Observation of Instrument Maintenance
 - Testing previously analyzed samples
 - Problem Solving

53

Six Levels of Competency

CLIA '88

All 6 levels must be assessed for each Test System a Tech is deemed Competent to perform. Must be re-evaluated if the method or instruments change.

- Direct Observation of Routine patient test performance
 - Patient identification and preparation
 - Specimen collection, handling, processing and testing
- Monitoring the recording and resulting of test results
 - Reporting of Critical Values
- Review of intermediate test results or worksheets
 - QC records, Proficiency Testing results and preventive maintenance records
- Direct Observation of instrument maintenance and function checks
- Assessment of test performance
 - Using previously analyzed specimens,
 - Blind testing samples
 - External Proficiency Testing samples
- Evaluation of problem solving skills

54

